



Running Contract Details	
Equipment Name	Tonometer Non contact
Running Contract Valid Till	13-12-2020
Tender Ref No	KMSCL/EP/T288/124/2018(R)
Tendered Quantity	51
Supplier Name	M/s Appasamy Associates
GST No	33AASCA3660R1Z0
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
# 20 - SBI Officers Colony 1st Street Arumbakkam Chennai-600106	Contact Person	B Radhakrishnan
	Phone	04430101400, 30101401
	Mobile No	09349211170
	Email	info@appasamy.com, appasamy.thrissur@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Tonometer Non contact</b> <i>Model &amp; Make : CANON-TX 20P / Canon singapore Pte.Ltd</i>	840000 Incl.GST :12%	61950	901950
		<b>840000</b>	<b>61950</b>	<b>901950</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Tonometer Non contact</b>							
Labour	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00	5,000.00
Comprehen sive	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00	20,000.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Tonometer Non contact**

1. Should have manual and automatic operation.
2. IOP measurement range should have 1mmHg to 60 mmHg
3. Should have minimum 11mm working distance.
4. Should have CCT correction function.
5. Should have fixation light (Blinking and non-blinking)
6. Chin rest vertical movement should have minimum  $\pm 30$  mm.
7. Should have minimum  $\pm 22$  mm forward /backward movement of measuring unit.
8. Should have minimum horizontal distance is  $\pm 43$  mm and vertical distance  $\pm 17$  mm of measuring unit.
9. Should have monitor for display.
10. Should have built in printer
11. Should be supplied motorised table with locking wheel.
12. Should be operated from 200 to 240 V/50Hz power supply.
13. Should be supplied 10 Nos paper roll for printer.
14. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.