



Running Contract Details	
Equipment Name	Traction Unit
Running Contract Valid Till	20-06-2021
Tender Ref No	KMSCL/EP/T322/126B/2019
Tendered Quantity	100
Supplier Name	M/s Bharat Medical System
GST No	33AEXPJ4035M1Z2
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
M-1074(New#15) - 7th Main Road Thrivalluvar Nagar Thiruvanniyur Chennai-600041	Contact Person	M Balasubramanian
	Phone	044-24516904/539
	Mobile No	9444052532
	Email	bmsmadras@gmail.com, bmschennai@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Traction Unit</b> <i>Model &amp; Make : Physio-Trac Plus (Digital) / Bharat Medical Systems</i>	29680 Incl.GST :12%	2188.9	31868.9
		<b>29680</b>	<b>2188.9</b>	<b>31868.9</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)			
Rate	3 <sup>rd</sup> Year	4 <sup>th</sup> Year	5 <sup>th</sup> Year
<b>Traction Unit</b>			
Labour	1,000.00	1,000.00	1,000.00
Comprehensive	1,500.00	1,500.00	1,500.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Traction Unit**

#### I. Traction Unit

1. Should be microprocessor controlled user friendly traction system.
2. Should have possibility for both cervical and lumbar traction.
3. Should have treatment time from 0 to 60 minutes
4. Should have at least three modes of operation.
5. Should have possibility for hold and rest time set.
6. Should have facility to set traction force from 5 to 45 kgs.
7. Should operate in mains supply 200 to 240 Vac, 50Hz.
8. Should be supplied with the following accessories.
  - a. Spreader bar
  - b. Patient stop switch
  - c. Cervical collar

#### II. TRACTION TABLE

1. Should have facility to fit any traction unit.
2. Should be a two section table.
3. Should have armpit for counter traction
4. Should have facility to adjust the height of the traction unit.
5. Should have powder coated mild steel body.
6. Should be supplied with soft mattress.
7. Should be supplied with foot rest.