

Running Contract Details	
Equipment Name	Transcutaneous Bilirubinometer
Running Contract Valid Till	08-08-2021
Tender Ref No	KMSCL/EP/T328/684/2019(R)
Tendered Quantity	100
Supplier Name	M/s Draeger Medical India Pvt Ltd
GST No	32AACCD3263D1Z9
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
35/230D - Surya Prabha Building Automobile Road Palarivattom Kochi - 682 025	Contact Person	
	Phone	0484 4013118
	Mobile No	9567862639
	Email	prem.kumar@draeger.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Transcutaneous Bilirubinometer</b> <i>Model &amp; Make : Draegor JM105 / Dragerwek AG &amp; Co KGaA</i>	359557.8 Incl.GST :18%	25169.05	384726.85
		<b>359557.8</b>	<b>25169.05</b>	<b>384726.85</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Transcutaneous Bilirubinometer</b>							
Labour	9,000.00	9,500.00	10,000.00	11,000.00	11,500.00	12,000.00	13,000.00
Comprehensive	65,000.00	70,000.00	80,000.00	90,000.00	1,00,000.00	1,10,000.00	1,20,000.00

### Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Transcutaneous Bilirubinometer**

1. Intended use : Pre, during and post phototherapy
2. Gestational age : 27-42 weeks
3. Post – natal age : 0-20 days
4. Patient weight range : 950-4995 grams
5. Total serum bilirubin range : 0-20 mg/dl / 0-340 $\mu$ mol/L
6. Accuracy (RMSE) : +/- 1.5mg/dL at 66% of the time or 1 Sigma  
+/- 26 $\mu$ mol/L at 66% of the time or sigma
7. Repeatability (SD) : +/-0.66 mg/dL  
+/-11.2  $\mu$ mol/L
8. Correlation : r=0.90
9. Expected battery life (minimum): 1 year
10. Should have safety certificate from a competent authority CE/FDA (US). Other equivalent certificates will not be accepted. Should have valid detailed electrical and functional safety test report from ERTL. Copy of the certificate/test report shall be produced along with the technical bid