



| Running Contract Details | |
|--------------------------------|--|
| Equipment Name | Transport incubator Model B |
| Running Contract Valid Till | 15-02-2021 |
| Tender Ref No | KMSCL/EP/T300/1317/2018 |
| Tendered Quantity | 16 |
| Supplier Name | M/s Nice Neotech Medical Systems Pvt Ltd |
| GST No | 33AABCN4311M1ZO |
| Installation & Delivery Period | 8 Week(s) |
| Up-time / PM vist | 95% & 4 Visits per year |
| Warranty period | 3 Years |

| Supplier`s Details | | |
|--|-----------------|-----------------------|
| Address | Contact Details | |
| No.85 Krishna Industrial Estate Vannagaram Mettukuppam Chennai 600 095 Tamil Nadu | Contact Person | K.Premkumar |
| | Phone | |
| | Mobile No | 9543455323 |
| | Email | sales@niceneotech.com |

| Item-wise Price Details | | | | |
|-------------------------|---|---|------------------------------------|-----------------|
| # | Item Details | Unit Rate (Incl.all taxes & charges) | Service Charges (Through KMSCL) | Grand Total |
| 1 | Transport incubator Model B <i>Model & Make : nice 3000 / M/s. Nice Neotech Medical Systems Pvt. ltd.</i> | 200480 Incl.GST :12% | 14785.4 | 215265.4 |
| | | 200480 | 14785.4 | 215265.4 |

| Annual / Comprehensive Maintenance Charges (Exl.Tax) | | | | | | | |
|--|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|
| Rate | 4 th Year | 5 th Year | 6 th Year | 7 th Year | 8 th Year | 9 th Year | 10 th Year |
| Transport incubator Model B | | | | | | | |
| Labour | 2,000.00 | 2,000.00 | 2,000.00 | 2,000.00 | 2,000.00 | 2,000.00 | 2,000.00 |
| Comprehensive | 2,500.00 | 2,500.00 | 2,500.00 | 2,500.00 | 2,500.00 | 2,500.00 | 2,500.00 |

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Transport incubator Model B

1. Working temperature range : 30 to 37 °C (Patient display mode) : 30 to 37°C (Air set mode).
2. Accuracy : +/- 0.2 °C, resolution : 0.1 °C
3. Accuracy of probe interchangeability : +/- 0.2°C
4. need for probe calibration : Not required
5. Temperature probe : Thermistor based interchangeable probe
6. Alarms :
 - i. High & Low temperature
 - ii. power failure
 - iii. Battery low
 - iv. Probe failure.
 - v. Fan failure
7. Voltage : 90 to 240 V at 50/60 Hz
8. Power :500 W maximum
9. Heating element : flat heaters.
10. Temperature display : bright numerical LED display
11. Messages and alarms : LCD display
12. Battery backup time : 2 hours
13. Performance specification :
 - i. Should have temperature rise time : 45 mins/10°C
 - ii. Should come with collapsible trolley for easy movement of equipment in to the ambulance.
 - iii. Should have in-built battery charger and battery should be part of equipment

- iv. Should have provision to attach 1 light weight oxygen cylinder and 1 light weight air cylinder (4.5 liter aluminium cylinders) with pin index mechanism
- v. Should have two elbow operatable ports.
- vi. Should have head access doors to pullout the bed to access the head of the baby during transportation
- vii. Should have front access door.
- viii. Should have one iris port and minimum 3 tubing port, 3 disposable infant restraint straps.
- ix. Should have an indicator for power ON
- x. Should have an indicator for heater ON
- xi. Equipment weight should be maximum 36 Kg (without cylinders)
- xii. Maximum size (L x W x H) cm should be with pole of 104 x 61 x 70 cm & without pole: 104 x 61 x 50 cm.
- xiii. Should have four wheels, at least two with breaks.
- xiv. Should have IV stand.
- xv. Should have goose neck examination lamp for flexible examination.
- xvi. Coating: epoxy/powder coated body for scratch and rust prevention.
- xvii. Should be supplied from a manufacturing company having ISO 9001:2008 and ISO 13485.
- xviii. Should have optional electrical suction
- xix. Should have an indicator for heater.