

| Running Contract Details       |                                     |
|--------------------------------|-------------------------------------|
| Equipment Name                 | Trop T/I Analyzer                   |
| Running Contract Valid Till    | 17-01-2021                          |
| Tender Ref No                  | KMSCL/EP/T294/239/2018              |
| Tendered Quantity              | 104                                 |
| Supplier Name                  | M/s Roche Diagnostics India Pvt Ltd |
| GST No                         | 33AAACA0266H1Z7                     |
| Installation & Delivery Period | 8 Week(s)                           |
| Up-time / PM vist              | 95% & 4 Visits per year             |
| Warranty period                | 3 Years                             |

| Supplier`s Details  |                 |                         |
|---|-----------------|-------------------------|
| Address   | Contact Details |                         |
| Sreyas Chamiers Tower<br>8th Floor West wing New No. 37<br>Old NO 23&24 - Chamiers Road<br>Teynampet<br>Chennai-600 018 | Contact Person  | Deepak Ranjan           |
|   | Phone           | 044-43900300            |
|   | Mobile No       | 919742746564            |
|   | Email           | t.manjunathan@roche.com |

| Item-wise Price Details |   |   |                                    |                  |
|-------------------------|---|---|------------------------------------|------------------|
| #                       | Item Details  | Unit Rate<br>(Incl.all taxes & charges) | Service Charges<br>(Through KMSCL) | Grand Total      |
| 1                       | <b>Trop T/I Analyzer</b><br><i>Model &amp; Make : Roche Cobas h232 / Roche Diagnostics GmbH</i> | 147500<br>Incl.GST :18%                 | 10325                              | 157825           |
| 2                       | <b>Test Cards</b>   | 459.2<br>Incl.GST :12%                  | 33.87                              | 493.07           |
|                         |   | <b>147959.2</b>                         | <b>10358.87</b>                    | <b>158318.07</b> |

| Annual / Comprehensive Maintenance Charges (Exl.Tax) |                      |                      |                      |                      |                      |                      |                       |
|--|----------------------|----------------------|----------------------|----------------------|----------------------|----------------------|-----------------------|
| Rate   | 4 <sup>th</sup> Year | 5 <sup>th</sup> Year | 6 <sup>th</sup> Year | 7 <sup>th</sup> Year | 8 <sup>th</sup> Year | 9 <sup>th</sup> Year | 10 <sup>th</sup> Year |
| <b>Trop T/I Analyzer</b>                             |                      |                      |                      |                      |                      |                      |                       |
| Labour   | 8,750.00             | 8,750.00             | 8,750.00             | 8,750.00             | 8,750.00             | 8,750.00             | 8,750.00              |
| Comprehensive  | 22,000.00            | 22,000.00            | 22,000.00            | 22,000.00            | 22,000.00            | 22,000.00            | 22,000.00             |

## **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

## **Technical Specification**

### **Equipment :Trop T/I Analyzer**

1. Should be able to measure high sensitive quantitative measurement of Troponin-T / Troponin I. Should be at least third generation assay.
2. It should be a portable unit and have inbuilt battery backup.
3. It should have parameter wise QC checking facility.
4. Should have a memory capacity of minimum 500 test results with date time and comments.
5. Should have minimum 200 code chip records (100 test+100QC).
6. Should have a minimum sample volume of approximately 300 micro litre.
7. Should have an operating temperature ranging from 18-30 degree Celsius.
8. Should work on 200-240Vac 50Hz power supply.
9. It should have an internal printing facility or supplied with an external thermal printer.
10. Should be supplied with 50 strips of Troponin T/I Parameter.
11. Should be provided with calibration certificate issued by the manufacturer at the time of installation.
12. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL.