



Running Contract Details	
Equipment Name	Tube sealer
Running Contract Valid Till	22-01-2020
Tender Ref No	KMSCL/EP/T248/128/2017(R)
Tendered Quantity	20
Supplier Name	M/s Kochin Inlab Equipments India Pvt Ltd
GST No	32AADCK3263F1ZX
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Door #28/2391-D Vihaan - Chilavannur Kadavanthra P.O Kochi-682020.	Contact Person	Manoj
	Phone	
	Mobile No	9447790938
	Email	manoj.inlab@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Tube sealer <i>Model & Make : XS 1010 / TERUMO PENPOL PVT LTD</i>	85417.92 Incl.GST :12%	6299.57	91717.49
		85417.92	6299.57	91717.49

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Tube sealer							
Labour	4,023.00	4,023.00	4,425.00	4,425.00	5,086.00	5,086.00	5,369.00
Comprehensive	5,900.00	5,900.00	6,490.00	6,490.00	7,139.00	7,139.00	7,906.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Tube sealer

1. Should be compatible with all ranges of blood bag systems available in the market
2. Should be heavy duty radio frequency sealer
3. Should be for bench-top use.
4. The sealing time should not be more than 2 seconds.
5. Sealing triggering should be automatic
6. Should have Automated tube detection
7. Should have Radiofrequency sealing with no risk of hemolysis of blood in tube segments
8. Should not be any risk of contamination.
9. Should have indication lamps detailing the functional status of sealer.
10. No warm-up time should be required.
11. Should ensure easy separation of tube segments after the sealing.
12. Should be operational on 220 to 240 V at 50 Hz