



Running Contract Details	
Equipment Name	Ultrasonic Cleaner
Running Contract Valid Till	18-06-2021
Tender Ref No	KMSCL/EP/T322/440A/2019
Tendered Quantity	200
Supplier Name	M/s Aczet Pvt Ltd
GST No	27AABCC3880G1Z5
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 2 Visits per year
Warranty period	2 Years

Supplier`s Details		
Address	Contact Details	
E-2 Plot No : 15 WICEL Opp. Seepz Gate No. 1 Andheri (East) Mumbai - 400 093	Contact Person	Mr. KARAN THAKUR
	Phone	022-42437748
	Mobile No	8452002414
	Email	tender@aczet.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Ultrasonic Cleaner <i>Model & Make : CUB 10 / ACZET PVT LTD</i>	32049.98 Incl.GST :18%	2243.5	34293.48
		32049.98	2243.5	34293.48

Annual / Comprehensive Maintenance Charges (Exl.Tax)				
Rate	3 rd Year	4 th Year	5 th Year	
Ultrasonic Cleaner				
Labour	1,375.00	1,375.00	1,375.00	
Comprehensive	1,800.00	1,800.00	1,800.00	

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Ultrasonic Cleaner

1. Should be able to remove the dirt and debris from the crevices of instruments.
2. Should have a SS 304 grade sturdy, non corrosive tank and casing.
3. Should have a tank capacity of approximately (6-8) Litre.
4. Should be supplied with standard accessories including SS 304 grade mesh bucket & lid.
5. Should have 42 KHz transducer (170 Watts or better)
6. Should have heating facility upto 60 degree.
7. Should have Electronic controller cum indicator with time set feature for heating and sonication, LED display also showing remaining time of cleaning.
8. Should have Water drain facility
9. Should operate on mains 220-240Vac, 50 Hz single phase.
10. Maximum sonic power and heating power need to be quoted.
11. Should be provided with a suitable voltage stabilizer of appropriate ratings