

Running Contract Details	
Equipment Name	Unweighing System Model B
Running Contract Valid Till	19-03-2021
Tender Ref No	KMSCL/EP/T311/685B/2019(R)
Tendered Quantity	15
Supplier Name	M/s Hospimedica International Ltd
GST No	07AAACH3234L1ZO
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
M/s Hospimedica International Ltd Leelawanti House 58/10 Ashok Nagar New Delhi-110018	Contact Person	Ashok Arora
	Phone	01125146165
	Mobile No	9895377669
	Email	mail@hospimedica.in

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Unweighing System Model B <i>Model & Make : NxStep Unweighing system / Biodex Medical System Inc, USA</i>	1734208 Incl.GST :12%	127897.84	1862105.84
		1734208	127897.84	1862105.84

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Unweighing System Model B							
Labour	30,968.00	30,968.00	30,968.00	30,968.00	30,968.00	30,968.00	30,968.00
Comprehensive	61,936.00	61,936.00	61,936.00	61,936.00	61,936.00	61,936.00	61,936.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Unweighing System Model B

1. The unit should have dynamic suspension system to maintain consistent unweighing during standing and walking.
2. The unit should have unique integral lift mechanism to assist patient from the seated position.
3. The unit should have an open frame design to accommodate all type of patients and to have an unobstructed forward view for the patients.
4. The unit should have display the readout quantifies unloaded weight.
5. The unit should have a hand held remote for height and unweighing load adjustment.
6. The unit should have open, unobstructed frame to allow clinician easy access to manually assist placement and timing of the lower extremities.
7. The unit should have lightweight, high strength aluminium frame to make it easy to move around.
8. The unit should have large, easy roll locking casters.
9. The unit should have removable and adjustable arm supports.
10. The unit should have one system that can accommodate children to adults.
11. The unit should provide up to 150 lb. (68 kg) unweighing capacity.
12. The unit should have a user capacity up to 170 kg.
13. The unit should have a vertical adjustment of 50 inches.
14. The unit should not weigh more than 130 kg.
15. The unit should have auto unload feature to compensate for movement due to slippage or posture changes.
16. The unit should have choice of support vests accommodates all size patients.
17. The unit should include universal support vest that accommodates chest size of 24 inches to 56 inches.
18. The unit should have international safety standards.
19. The unit should be able to reduce width by adjusting a lever.