



Running Contract Details	
Equipment Name	Vaccum extractor
Running Contract Valid Till	15-07-2020
Tender Ref No	KMSCL/EP/T267/327/2017(R)
Tendered Quantity	250
Supplier Name	M/s Kerala Surgical Equipment Co
GST No	32AACFK7295D1ZG
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
No.41/1737 A MES Centre Near Town Hall Banerji Road Cochin -18	Contact Person	Mr. Abbas
	Phone	
	Mobile No	9539012460
	Email	info@keralasurgicals.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Vaccum extractor <i>Model & Make : Digital LED Model MVDS-02 / MEDISIL</i>	144144 Incl.GST :12%	10630.62	154774.62
		144144	10630.62	154774.62

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Vaccum extractor							
Labour	3,068.00	3,068.00	3,068.00	3,068.00	3,068.00	3,068.00	3,068.00
Comprehensive	7,670.00	7,670.00	7,670.00	7,670.00	7,670.00	7,670.00	7,670.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Vaccum extractor

1. Should have Vacuum generator , silastic cup and metal cup of different sizes with necessary electrical connection cable.
2. Should have Foot Switch operation.
3. Should generate a steady vacuum of at least 600mmHg.
4. Should work on 200-240Vac 50Hz power supply.
5. Should have digital display.