

Running Contract Details	
Equipment Name	Video Colposcope
Running Contract Valid Till	30-07-2020
Tender Ref No	KMSCL/EP/T264/909/2017
Tendered Quantity	60
Supplier Name	M/s Sree Gokulam Healthcare Pvt Ltd
GST No	32AAJCS3502K1ZI
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Dhanya Complex Sastha Temple Road Ernakulam North PO Cochin-682018	Contact Person	Anoop
	Phone	
	Mobile No	9995813333
	Email	sales@gokulamhealthcare.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	Video Colposcope <i>Model & Make : C6A / Edan Instruments Inc</i>	221480 Incl.GST :12%	16334.15	237814.15
		221480	16334.15	237814.15

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 th Year	5 th Year	6 th Year	7 th Year	8 th Year	9 th Year	10 th Year
Video Colposcope							
Labour	11,800.00	11,800.00	11,800.00	11,800.00	11,800.00	11,800.00	11,800.00
Comprehensive	35,400.00	35,280.00	39,028.50	41,005.00	43,070.00	45,194.00	47,436.00

Other terms & conditions

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender

document).

2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

Technical Specification

Equipment :Video Colposcope

1. Should have variable light intensity and green filter.
2. Manual/Auto focus should be available
3. Facility to connect DESKTOP PC & LAPTOP PC with the system
4. Should have video camera with colour CCD with resolution of minimum 795 x 596 and TV resolution 480 lines or above.
5. Should have video output
6. The camera should have an in built light source with ultra-bright white LED, circular cool light source group with two or more step magnification.
7. Should have a camera work distance between 200 to 300mm.
8. The camera should have image freeze, Timer and Electronic Green filter, focusing with "1-30X" auto or manual focus mode
9. Should have an average lamp life of > 20000 hours.
10. Should have a PC based workstation for image data acquisition, saved image can be edited into diagnostic reports and analyzed.
11. Should have the ability to record live Video of cervical examination and surgical treatment procedures with Video image management, image review and annotation, customized reports printing, Reid Index, Biopsy marking, DVD/CD burn and electronic transfer of images/reports by email.
12. The camera should have 2 or more step magnification magnification.
13. Computer (i3, 8GB, 1TB, 19'' monitor with ups) / PC laptop (display 15'') and printer should be provided.
14. Powder coated MS trolley shall be supplied along with the machine