

Running Contract Details	
Equipment Name	Wheel Chair
Running Contract Valid Till	24-10-2020
Tender Ref No	KMSCL/EP/T277/361A/2018
Tendered Quantity	5000
Supplier Name	M/s Medimek Industries
GST No	27ADBPC1957E1ZY
Installation & Delivery Period	6 Week(s)
Up-time / PM vist	95% & 0 Visits per year
Warranty period	1 Years

Supplier`s Details		
Address	Contact Details	
B-50 Additional MIDC Anand Nagar Ambernath(East) - 451506 Thane Maharashtra	Contact Person	Jobin Augustin
	Phone	
	Mobile No	8281815893
	Email	medimek@ymail.com, medimek@rediffmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Wheel Chair</b> <i>Model &amp; Make : Mi-6077 / Medimek Industries</i>	15523.2 Incl.GST :5%	1221.16	16744.36
		<b>15523.2</b>	<b>1221.16</b>	<b>16744.36</b>

### **Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).

5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment :Wheel Chair**

1. Size : 790 mm (L) x 600mm (W) x 780 (H)
2. Structure : made of 22x1.2 mm SS 304 grade
3. Finish : Chrome finish
4. Crossbar : 25.4x 1.2mm thick
5. Material : S.S Tubular framework fitted with S.S Seat and back of grade SS 304.
6. Wheels : Two solid PU wheels and self-propelling S.S Loops. Two swivel castors 100 mm diameter in front.
7. Arm rest : ABS plastic
8. Foot rest : Aluminium die cast
9. Should have cushion for seat.
10. Should comply with National/International standards (CE/FDA)