



Running Contract Details	
Equipment Name	Whole body phototherapy chamber
Running Contract Valid Till	13-12-2020
Tender Ref No	KMSCL/EP/T275/433/2018
Tendered Quantity	50
Supplier Name	M/s Universal Agencies
GST No	32AAAFU5662Q1ZQ
Installation & Delivery Period	8 Week(s)
Up-time / PM vist	95% & 4 Visits per year
Warranty period	3 Years

Supplier`s Details		
Address	Contact Details	
Palakkal Angadi Thrissur-680001	Contact Person	Mr. Renjith Menon, Mr.Manoj kumar
	Phone	0487-2440319, 0487-2427031
	Mobile No	9072844272, 9447053244
	Email	uniage.sales@gmail.com

Item-wise Price Details				
#	Item Details	Unit Rate (Incl.all taxes & charges)	Service Charges (Through KMSCL)	Grand Total
1	<b>Whole body phototherapy chamber</b> <i>Model &amp; Make : SQUARE 12+12 ANB / HEALKEREN MEDTRONICS PVT LTD</i>	307440 <small>Incl.GST :12%</small>	22673.7	330113.7
		<b>307440</b>	<b>22673.7</b>	<b>330113.7</b>

Annual / Comprehensive Maintenance Charges (Exl.Tax)							
Rate	4 <sup>th</sup> Year	5 <sup>th</sup> Year	6 <sup>th</sup> Year	7 <sup>th</sup> Year	8 <sup>th</sup> Year	9 <sup>th</sup> Year	10 <sup>th</sup> Year
<b>Whole body phototherapy chamber</b>							
Labour	3,000.00	4,000.00	5,000.00	6,000.00	7,000.00	8,000.00	9,000.00
Comprehen sive	10,000.00	12,000.00	14,000.00	18,000.00	20,000.00	22,000.00	24,000.00

**Other terms & conditions**

1. The supplier shall execute an agreement with the purchaser as per tender conditions (agreement format is given in the tender document).
2. The supplier shall submit performance security amounting to 5% of the value of the supply order.
3. The labour & comprehensive charges of equipment after the completion of warranty period is finalized by KMSCL as mentioned above.
4. Since discount rate is not applicable for equipment under Running Contract of KMSCL, purchase/supply order can be issued directly to supplier at the given rate with tax & other charges (exclusive of KMSCL service charges).
5. If purchase/supply order is issued directly to the supplier, KMSCL service charge need not be paid. But the copy of the said order may be forwarded to KMSCL for information.

### **Technical Specification**

#### **Equipment : Whole body phototherapy chamber**

1. Dimensions Outer (mm) : 1955(H) x 1070(W) x 1070(D)  $\pm$  20 mm.
2. Dimensions Inner (mm) : 1870(H) x 750(W) x 750(D)  $\pm$  20 mm.
3. Power Required : 6KVA (24tubes)
4. Tubes : TL 100W/10R
5. Reflectors : Imported aluminum mirror type
6. Painting : High Quality(7 line process) powder coating
7. Chokes : Special UV chokes
8. Capacity : approx 150kg
9. It should be provided with ELCB
10. The unit should have UVA & NBUVB combination of tubes
11. Special UV chokes for maximum life for the tubes
12. The unit is provided with mirror type reflectors
13. Cabinet should be made of high quality steel and should be powder coated.
14. Should have in-built sensor dosimeter with Dynamic Range Angular sensitivity(DRAS)
15. Cumulative hour meter should be provided with password.
16. Should have four line liquid crystal displays for control panel.
17. Should have weather touch key pad
18. Should have lock and key provided for the control panel.
19. Unit should function with 200-240V AC, 50/60 Hz input power supply
20. Should have safety certificate from a competent authority CE issued by a notified body registered in European Commission / FDA (US) / STQC CB certificate / STQC S certificate or valid detailed electrical and functional safety test report from ERTL. Copy of the certificate / test report shall be produced along with the technical bid.

